



METERING SALES & SERVICE, INC.

Distributors of Metering Equipment & Systems

1930 PIONEER ROAD
HUNTINGDON VALLEY, PA 19006
(215) 675-5117 • FAX (215) 675-4960

October 19, 2011

Ryan Marlborough
Photovoltaic Specialist
United Electric Supply Co.
10 Bellecor Drive
New Castle, DE 19720

Dear Mr. Marlborough:

The reconditioned Sangamo meter, S/N 41-198-493, that we shipped was fully tested and met the standards common to utilities. This meter was physically examined, cleaned and then tested utilizing a Watthour Engineering computerized test set with a Radian Research RM-11 Standard. When this meter left our facility, it measured +/- 1/2% full load and light load and +/- 1% power factor. These are the normally required test points.

If you have any further questions, please contact our office.

Very truly yours,


Steven J. Helzner

SJH:mms